CERTIFICATION

ANNIS MODEL 25 MAGNETIC FIELD INDICATORS

	Gauss Range	Temperature	F
It is hereby certified that the above se calibrated to meet manufacturer's new ANSI Z540.3 using a precision Helmh directly calculable, to a high degree of H=0.8992 NI R cm.	w instruments specification. Such noltz type uniform magnetic field so	testing being performed in conform testing being performed in which the magnetic field leads	nance with
These tests are conducted using Helmultiple laboratory type DC ammeters calibration traceable to N.I.S.T. The last and ards similar to those we manufall U.S. Air Force, United Testing Labs, (s two of which are cross-checked of Helmholtz Coil is cross-checked ag acture and supply for use in other o	on a daily basis and all of which ha gainst other certified Helmholtz ma alibration laboratories such as: U.	ave Ignetic field
These Model 25 Pocket Magnetometomoving-magnet type gauss meter that calibrated to be accurate within 2% of be within 5% of full scale.	at is commercially available. Instru	ments having ranges of 2 to 100 g	auss are
All moving magnet type gauss meters Model 25) is spaced 5/8" away from t	-	at the movement staff location wh	ich, (on
Earth's field bias must be eliminated when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer	nter zero, and when calibrating, that	
when the instrument pointer is aimed	magnetic west, it is exactly on cer	nter zero, and when calibrating, that	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer ith the dip angle of the local magne	nter zero, and when calibrating, thatetic field.	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer ith the dip angle of the local magne	nter zero, and when calibrating, thatetic field.	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer ith the dip angle of the local magne	nter zero, and when calibrating, thatetic field.	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer ith the dip angle of the local magne	nter zero, and when calibrating, thatetic field.	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer ith the dip angle of the local magne	nter zero, and when calibrating, thatetic field.	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	magnetic west, it is exactly on cer ith the dip angle of the local magne	nter zero, and when calibrating, thatetic field.	at the
when the instrument pointer is aimed instrument staff is oriented parallel wi	*AS RECEIVED *anufactured by, and is calibrated by	nter zero, and when calibrating, the etic field. FINAL CALIBRA	TION
when the instrument pointer is aimed instrument staff is oriented parallel wi STANDARD STANDARD This Model 25 Magnetometer was ma	*AS RECEIVED *AS RECEIVED anufactured by, and is calibrated by ument.	nter zero, and when calibrating, the etic field. FINAL CALIBRA	TION

^{*}This column is only used for Magnetometers sent in for re-calibration. This column will have dashes instead of values for Magnetometers purchased new.